



Certificate of Calibration

CCIS Item Code: 7326STX1

Certification No.: ***

Mfg. / Serial No.: *****

Range: 0.0000 – 0.0500"

Accuracy: +/- 5%

This gauge was certified using test equipment whose accuracy is traceable to the National Institute of Standards and Technology. We certify this gauge has been tested and falls within the manufacture's accuracy as noted.

Standard	Actual Reading	Allowable Range
No Shim	****	0.0000"
0.020"	****	.019 - .021
0.030"	****	.0285 - .0315
0.040"	****	.038 - .042

CALIBRATION STANDARD: Starrett 467 Thickness Gauge SERIAL NO. *****

DATE CALIBRATED: *****

CALIBRATION DUE DATE: *****

RANGE: .020-.040

The above standard is traceable for accuracy to the "National Institute of Standards and Technology" Test Plates

Cross Country Infrastructure Services